

OFFICE OF THE CHIEF MEDICAL AND HEALTH OFFICER
DISTRICT HEALTH SOCIETY, DISTRICT N.C.D CELL BILASPUR

Near Vedic convent school, Nutan chowk, sarkanda Bilaspur C.G.-492001

Phone No. 07752-424101

Email- ncdcellbsp@gmail.com

क्रमांक / एन.सी.डी / 2014 / 5559

दिनांक 11/07/2014


एन.सी.डी. कार्यक्रम अंतर्गत आमंत्रित निविदाओं को दिनांक 09/07/2014 को समय 12:00 बजे अध्यक्ष महोदय की अध्यक्षता में कवर । तकनीकी मूल्यांकन हेतु खोला गया। सरल क्रमांक 02 में उल्लेखित पात्र फर्मों को निम्नलिखित दस्तावेज कार्यालयीन समय में दिनांक 14/07/2014 समय 5:30 बजे तक अनिवार्य रूप से अधोहस्तारकर्ता कार्यालय में तथा ई-मेल ncdcellbsp.equipment@gmail.com पर जमा कराया जाना है, उक्त दिनांक तक जमा न होने की स्थिति में स्वतः निरस्त माना जावेगा। कवर B दिनांक 15/07/2014 समय 12:00 बजे स्थान अध्यक्ष महोदय (एनसीडी क्रय समिति) के कक्ष (कलेक्टर परिसर) में Price Bid खोला जावेगा।

List of Firm Rejected in Technical Bid (Cover -A)			
S.No	Document No.	Name of the Bidder	Reason of Rejection
1	4697	M/s Triveni Medi Sales	Average Annual Turnover is less than fifty (50) lacs as per 2.2 E.C

List of Firm qualified suomotto subject to submission of following documents before opening of Price-Bid (Cover-B)			
Last Date of submission of Documents		14/07/2014 Time : 5.30 PM	
S.No	Document No.	Name of the Bidder	Documents Required to be submitted
1	4695	M/s Nihon Kohden India Pvt Ltd	RTGS Receipt and Tender Document with seal and sign
2	4691	M/s Goyal Medico	Upto date Vat/sale tax clearance cetificate
3	4655	M/s Oswal Novelties	Manufacturing Licence/FDA Certificate/Import Licence and Turn over Certificate(without any over written on it)
4	4655	M/s Dee Dee Enterprise	Manufacturing Licence/FDA Certificate/Import Licence and Tender Document with Seal and sign
5	4689	M/s Vision Enterprise	Non-Conviction Affidavit
6	4693	M/s A R Corporation	Non- Black listed Affidavit of the Items Quoted
7	4698	M/s Anu sales Corporation	Manufacturing Licence/FDA Certificate/Import Licence

List of Products Diqualified/Rejected on the basis of technical evaluation in Technical Bid (Cover -A)

S.No	Name of the Bidder	Product Name	Reason for Rejection
1	M/s Scientific Traders	Cogulometer	Due to single channel measuring less than the desired technical specification
2	M/s Dee Dee Enterprise	Colposcope & Non-Invasive Ventilator	EMD not submitted and also not attached technical specification brochure
3	M/s Moorarka Agencies	TENS	Only 2 channel specified less than desired specification
4	M/s A R Corporation	Walker Cervical Traction with Kit Bags Multi channel Monitor ECG Machine Non-Invasive Ventilator Electrolyte Analyser	Due to less than the desired technical specification Voltage, Mode, force and other units not specified in the brochure Due to display size and Non-Modular Design less than the desired technical specification Due to non availability of Alphanumeric keyboard and VGA/USB Port Due to display size less than the desired technical specification Due to sample reading demanded for whole Blood sample as per the desired technical specification
5	M/s Bio Medics	ECG Machine	Due to Display size less than desired technical specification
6	M/s Sonu Medical Agency	ECG Machine Urine Analyser Centrifuge Machine Cogulometer	Due to Display size and pape size less than desired technical specification Due to non-performing of Ascorbic acid test Not mention of sample capillary holder no. of tables roated on the product brochure Measuring channel is less than the desired technical specification
7	M/s Anu sales Corporation	Urine Analyser	Performing Test not mention in the brochure
8	M/s Goyal Medico	ECG Machine Cogulometer	Alpha numeric keyboard and A4 size Thermal paper recorder not available Measuring channel is less than the desired technical specification


Chief Medical and Health officer
 Bilaspur (Chhattisgarh)